

ROMANIA
MINISTRY OF HEALTH
PROJECT MANAGEMENT UNIT

Health Sector Reform – Improving Health System Quality and Efficiency Project
IBRD Loan no. 8362-RO

ADDENDUM NO. 4 to the BIDDING DOCUMENTS
FOR
PROCUREMENT OF DIAGNOSTIC IMAGING MEDICAL EQUIPMENT - ICB No. G/C1/2.1

According with the *ITB 8 – Amendment of the Bidding Document*, the Bidding Documents is hereby amended as follows:

Section III. Evaluation and Qualification Criteria

Article 1. The following phrase under 3.1(i) Post-qualification Requirements (ITB 36.1) under Section III – Evaluation and Qualification Criteria is hereby modified and it will be read as follows:

“[...] The Bidder will also provide the Acceptance Certificates/ Delivery Protocol or equivalent (“*copy conform with the original*”) **signed for conformity by the authorized representative of the Bidder** to demonstrate the successful completion of the respective contracts. [...]”

Article 2. Requirement (v) of 3.1 Post-qualification Requirements (ITB 36.1) under Section III – Evaluation and Qualification Criteria is hereby modified and it will be read as follows:

“The Bidders must not have unpaid impositions or taxes towards the Romanian state budget or towards the state budget of their residency country (in case of foreign bidders).

The Bidder must submit a **certificate** (original, including online version/ legalized copy) or **equivalent document** issued by the National Agency of Fiscal Administration office or a similar entity (in case of foreign bidders) attesting that the bidder has no debts towards the state budget (Romanian or country of residency), valid on the bid submission date.”

The equivalent document can be in any form as long as it is signed for conformity by authorized person and accompanied by solemn declaration of authorized person indicating how the information can be verified with the issuing entity (e.g. website address etc.).”

Section VII. Schedule of Requirements

3. Technical Specifications

LOT II – MAGNETIC RESONANCE IMAGING EQUIPMENT

Article 3. The requirement “*Guaranteed minimum magnetic field homogeneity inside a spherical volume 45 cm in diameter: maximum 5,1 ppm*” under Lot II Magnetic resonance imaging equipment of chapter 3. Technical Specifications, Section VII – Schedule of Requirements, is hereby modified and it will be read as follows:

“Guaranteed minimum magnetic field homogeneity inside a **spheroidal volume of 45 - 50 cm in diameter: maximum 5,1 ppm**”



Handwritten signature in blue ink.

Article 4. The requirements of “**APPLICATIONS**” under Lot II Magnetic resonance imaging equipment of chapter 3. Technical Specifications, Section VII – Schedule of Requirements, are hereby completed with the following requirements:

“Cardiac imaging:

- Morphological and functional cardiac exploration;
- Cine phase-contrast velocity;
- Myocardial perfusion and late enhancement imaging;
- Coronary MR angiography”

Article 5. The requirements of “**Software applications available on the post-processing console**” under Lot II Magnetic resonance imaging equipment of chapter 3. Technical Specifications, Section VII – Schedule of Requirements, are hereby completed with the following requirements:

“Software for cardiac imaging post-processing”

Article 6. The requirement “Antenna used for abdominal/ pelvic examinations – minimum 12 channels” of “**Antennae**” under Lot II Magnetic resonance imaging equipment of chapter 3. Technical Specifications, Section VII – Schedule of Requirements, is hereby modified and it will be read as follows: with the following requirements:

“Antenna used for abdominal/ pelvic/ **cardiac** examinations – minimum 12 channels”

Article 7. All other parts, clauses and provisions of the original Bidding Documents that are not amended as per the articles above shall remain valid and unchanged.



Yelosi